



1. ELECTRICAL SPECIFICATIONS

Accuracy calculated as $\pm[\% \text{reading} + (\text{num dgt} * \text{resolution})]$ ta $18^{\circ}\text{C} \div 28^{\circ}\text{C}$, $<75\% \text{HR}$

DC VOLTAGE


Range (*)	Resolution	Accuracy	Input impedance	Overload protection
600.0mV	0.1mV	$\pm(0.9\% \text{rdg} + 5 \text{dgt})$	$>10 \text{M}\Omega$	1000VDC/ACrms
6.000V	0.001V			
60.00V	0.01V			
600.0V	0.1V			
1000V	1V	$\pm(1.2\% \text{rdg} + 5 \text{dgt})$		

(*) HIREs feature not active. With HIREs feature active, multiply x10 the dgts of the accuracy

AC TRMS VOLTAGE

Range (*)	Resolution	Accuracy (**)		Overload protection
		50Hz \div 1kHz	1kHz \div 5kHz	
600.0mV	0.1mV	$\pm(1.0\% \text{rdg} + 5 \text{dgt})$	$\pm(3.0\% \text{rdg} + 5 \text{dgt})$	1000VDC/ACrms
6.000V	0.001V			
60.00V	0.01V			
600.0V	0.1V			
1000V	1V		Not specified	

(*) HIREs feature not active. With HIREs feature active, multiply x10 the dgts of the accuracy

(**) Accuracy specified from 5% to 100% of the measuring range and crest factors ≤ 3 , Input impedance: $> 10 \text{M}\Omega$ Accuracy PEAK function: $\pm(10\% \text{rdg} + 30 \text{dgt})$, Response time PEAK function: 1msAccuracy  function: $\pm(1.5\% \text{rdg} + 5 \text{dgt})$ ($f \leq 60 \text{Hz}$), $\pm(3.0\% \text{rdg} + 5 \text{dgt})$ ($60 \text{Hz} \leq f \leq 1 \text{kHz}$), not specified ($f: 1 \div 5 \text{kHz}$)For not sinusoidal waveforms the accuracy is: $\pm(4.5\% \text{rdg} + 10 \text{dgt})$ ($50 \text{Hz} \div 1 \text{kHz}$), $\pm(10.0\% \text{rdg} + 10 \text{dgt})$ ($1 \text{kHz} \div 5 \text{kHz}$)

AC+DC TRMS VOLTAGE

Range (*)	Resolution	Accuracy (50Hz \div 1kHz)	Input impedance	Overload protection
600.0mV (*)	0.1mV	$\pm(1.5\% \text{rdg} + 10 \text{dgt})$	$>10 \text{M}\Omega$	1000VDC/ACrms
6.000V	0.001V			
60.00V	0.01V			
600.0V	0.1V			
1000V	1V			

(*) HIREs feature not active. With HIREs feature active, multiply x10 the dgts of the accuracy


DC/AC TRMS VOLTAGE WITH LOW IMPEDANCE (LoZ)

Range (*)	Resolution	Accuracy (**)	Input impedance	Overload protection
6.000V	0.001V	$\pm(3.0\% \text{rdg} + 40 \text{dgt})$	ca $3 \text{k}\Omega$	1000VDC/ACrms
60.00V	0.01V			
600.0V	0.1V			
1000V	1V			

(*) HIREs feature not active. With HIREs feature active, multiply x10 the dgts of the accuracy

(**) AC+DC accuracy: $\pm(3.5\% \text{rdg} + 40 \text{dgt})$

DIODE TEST

Range	Max test current	Open voltage
	$<1.5 \text{mA}$	2.8V

4-20mA% READINGS

Range	Resolution	Accuracy	Readings
$-25\% \div 125\%$	0.1%	$\pm 50 \text{dgt}$	0mA=-25%, 4mA=0%, 20mA=100%, 24mA=125%



DC CURRENT

Range (*)	Resolution	Accuracy	Overload protection
600.0μA	0.1μA	±(1.5%rdg+5dgt)	Fast Fuse 0.8A/1kVAC/DC (inputs mA, μA)
6000μA	1μA		
60.00mA	0.01mA		
600.0mA	0.1mA		
10.00A	0.01A		Fast Fuse 10A/1kVAC/DC (input 10A)

(*) HIRES feature not active. With HIRES feature active, multiply x10 the dgts of the accuracy

AC TRMS CURRENT

Range (*)	Resolution	Accuracy (**)		Overload protection
		50Hz ÷ 1kHz	1kHz ÷ 5kHz	
600.0μA	0.1μA	±(2.5%rdg+5dgt)	±(3.5%rdg+5dgt)	Fast Fuse 0.8A/1kVAC/DC (inputs mA, μA)
6000μA	1μA			
60.00mA	0.01mA			
600.0mA	0.1mA			
10.00A	0.01A			Fast Fuse 10A/1kVAC/DC (input 10A)

(*) HIRES feature not active. With HIRES feature active, multiply x10 the dgts of the accuracy

(**) Accuracy specified from 5% to 100% of the measuring range, Accuracy PEAK function: ±(10%rdg + 30dgt)

For not sinusoidal waveforms the accuracy is: ±(4.5%rdg + 10dgt) (50Hz÷1kHz), ±(10.0%rdg + 10dgt) (1kHz÷5kHz)

Accuracy AC+DC: ±(2.8%rdg+5dgt) (50Hz ÷ 1kHz)

DC CURRENT WITH TRANSDUCER CLAMPS

Range (*)	Output ratio	Resolution	Accuracy (**)	Overload protection	
1000mA	1000mV/1000mA	1mA	±(1.5%rdg + 6dgt)	1000VDC/ACrms	
10A	100mV/1A	0.01A			
30A					
40A					
100A	10mV/1A	0.1A			±(1.5%rdg + 26dgt)
300A					±(1.5%rdg + 6dgt)
400A	1mV/1A	1A			±(1.5%rdg + 26dgt)
1000A					±(1.5%rdg + 6dgt)

(*) HIRES feature not active. With HIRES feature active, multiply x10 the dgts of the accuracy

(**) Accuracy referred to only instrument without transducer clamp

AC TRMS CURRENT WITH TRANSDUCER CLAMPS

Range (*)	Output ratio	Resolution	Accuracy (**, ***)	Overload protection	
1000mA	1000mV/1000mA	1mA	±(2.5%rdg + 10dgt)	1000VDC/ACrms	
10A	100mV/1A	0.01A			
30A					
40A					
100A	10mV/1A	0.1A			±(3.5%rdg + 30dgt)
300A					±(2.5%rdg + 10dgt)
400A	1mV/1A	1A			±(2.5%rdg + 10dgt)
1000A					±(3.5%rdg + 30dgt)
3000A			±(2.5%rdgh + 10dgt)		

(*) HIRES feature not active. With HIRES feature active, multiply x10 the dgts of the accuracy

(**) Accuracy referred to only instrument without transducer clamp

(***) Accuracy specified from 5% to 100% of the measuring range; For pulse waveforms the accuracy is: ±(10.0%rdg + 10dgt)

Accuracy PEAK function: ±(10%rdg + 30dgt), AC+DC TRMS current: accuracy (50Hz÷1kHz): ±(3.0%rdg + 10dgt)

For not sinusoidal waveforms the accuracy is: ±(4.5%rdg + 10dgt) (50Hz÷1kHz)

**RESISTANCE AND CONTINUITY TEST**

Range (*)	Resolution	Accuracy	Buzzer	Overload protection
600.0Ω	0.1Ω	±(2.0%rdg+9dgt)	<35Ω	1000VDC/ACrms
6.000kΩ	0.001kΩ	±(1.2%rdg+5dgt)		
60.00kΩ	0.01kΩ			
600.0kΩ	0.1kΩ			
6.000MΩ	0.001MΩ	±(2.0%rdg+10dgt)		
60.00MΩ	0.01MΩ			

(*) HIRES feature not active. With HIRES feature active, multiply x10 the dgts of the accuracy

FREQUENCY (Electrical circuits)

Range (*)	Resolution	Accuracy	Sensitivity	Overload protection
40.00Hz ÷ 10.00kHz	0.01Hz	±(0.5%rdg)	15Vrms	1000VDC/ACrms

(*) HIRES feature not active. With HIRES feature active, multiply x10 the dgts of the accuracy

FREQUENCY (Electronic circuits)

Range (*)	Resolution	Accuracy	Sensitivity	Overload protection
60.000Hz	0.001Hz	±(1.0%rdg+2dgt)	>2.0Vrms min (20% < duty < 80%, <100kHz) 5Vrms min (20% < duty < 80%, >100kHz)	1000VDC/ACrms
600.00Hz	0.01Hz			
6.0000kHz	0.0001kHz			
60.000kHz	0.001kHz			
600.00kHz	0.01kHz			
6.0000MHz	0.0001MHz			
10.000MHz	0.001MHz			

(*) HIRES feature not active. With HIRES feature active, multiply x10 the dgts of the accuracy

DUTY CYCLE

Range (*)	Resolution	Accuracy	Overload protection
0.1 ÷ 99.9%	0.01%	±(1.2%rdg+2dgt)	1000VDC/ACrms

(*) HIRES feature not active. With HIRES feature active, multiply x10 the dgts of the accuracy

Pulse width: 100μs ÷ 100ms ; Frequency: 5Hz ÷ 100kHz

CAPACITANCE

Range	Resolution	Accuracy	Overload protection
60.00nF	0.01nF	±(3.5%rdg+10dgt)	1000VDC/ACrms
600.0nF	0.1nF		
6.000μF	0.001μF	±(2.5%rdg+10dgt)	
60.00μF	0.01μF		
600.0μF	0.1μF		
6.000mF	0.001mF	±(3.5%rdg+10dgt)	

TEMPERATURE WITH TYPE K PROBE

Range	Resolution	Accuracy (*)	Overload protection
-40.0°C ÷ 600.0°C	0.1°C	±(2.0%rdg+3°C)	1000VDC/ACrms
600°C ÷ 760°C	1°C		
-58.0°F ÷ 600.0°F	0.1°F	±(2.0%rdg+5.5°F)	
600°F ÷ 1400°F	1°F		

(*) Accuracy referred to instrument without probe



2. GENERAL SPECIFICATIONS

Display:

- LCD, 3 ½ dgt (6000counts) and 4½dgt (60000 counts) decimal point and bargraph
- Automatic polarity indication
- Backlight
- "OL" over range indication
- Response time: 3/s
- Conversion: TRMS

Features:

- Data HOLD, Auto HOLD
- MAX/MIN
- PEAK (Voltage and Current)
- RANGE
- REL
- DC 4-20mA% current readings

Power supply:

- 4 x 1.5V alkaline batteries type AAA LR03
- Battery life: ca 18h (backlight ON), ca 60h (backlight OFF)
- Auto Power OFF after 15 minutes of idleness

Mechanical specifications

- Dimensions (L x W x H): 175 x 85 x 55mm
- Weight (included batteries): 360g
- Mechanical protection: IP40

Environmental conditions:

- Working temperature: 5°C ÷ 40°C
- Working humidity: <80%RH
- Storage temperature: -20°C ÷ 60°C
- Storage humidity: <80%RH
- Altitude max of use: 2000m

Reference guidelines:

- Safety : IEC/EN61010-1
- EMC : IEC/EN61326-1
- Pollution degree: 2
- Insulation: double insulation
- Measurement category: CAT IV 600V – CAT III 1000V

This product conforms to the prescriptions of the European directive on low voltage 2014/35/EU and to EMC directive 2014/30/EU

This product conforms to the prescriptions of the European directive 2011/65/EU (RoHS) and the European directive 2012/19/EU (WEEE)

Services d'EURO-INDEX

EURO-INDEX est un fabricant, importateur et distributeur de diverses marques A dans le domaine des instruments de test et de mesure. Nous fournissons également une large gamme de services pour optimiser l'utilisation de ces instruments dans vos activités. Ces services comprennent naturellement l'entretien, la réparation et l'étalonnage des instruments, mais nous proposons aussi une assistance sous forme de formation via notre EURO-INDEX Academy et la location d'instruments.

Centre de Service Agréé

EURO-INDEX est un Centre de Service Agréé pour toutes les marques représentées. Cela signifie que vos instruments sont pris en charge par des techniciens formés par le fabricant et disposant des outils et logiciels adéquats. Seules des pièces d'origine sont utilisées et la garantie de votre instrument, ainsi que les certifications (ATEX, EN50379, etc.) restent intactes.

Laboratoire de maintenance et de calibrage

Le laboratoire des Pays-Bas est accrédité RvA selon la norme EN-ISO/IEC 17025. Cette accréditation est valable pour différentes grandeurs, telles que spécifiées dans le champ d'application associé au numéro d'accréditation K105. Les certificats de calibrage RvA sont acceptés à l'international et équivalents à ceux de BELAC.



Service Mobile

Outre les laboratoires d'étalonnage fixes de Zaventem et de Capelle aan den IJssel, nous disposons également d'un laboratoire itinérant appelé "Service mobile". Nos services peuvent venir vers vous, en offrant une qualité équivalente.

MQS®

MQS® est une formule d'entretien exclusive comportant un entretien et un calibrage périodiques de vos instruments de mesure à un coût fixe et faible. Via un portail Web gratuit (monmq.be), vous avez toujours accès à vos certificats de calibrage.

Location d'instruments de mesure

- Vaste assortiment
- Précision démontrable par le certificat d'étalonnage actuel
- Conseils avisés
- Les instruments sont livrés avec leurs accessoires

EURO-INDEX Academy

- Formations et séminaires
- Vidéos de démonstration et d'instruction
- Notes d'application



Comptoir de service



Entretien, réparation et calibrage



Formations et séminaires



Service Mobile

Sous réserve de modifications EURO-INDEX® FR 23001



BELGIQUE
Chaussée de Louvain 607
1930 Zaventem
T: 02 - 757 92 44
F: 02 - 757 92 64
sales@euro-index.be
www.euro-index.be

PAYS-BAS
Rivium 2e straat 12
2909 LG Capelle a/d IJssel
T: +31 - (0)10 - 2 888 000
F: +31 - (0)10 - 2 888 010
verkoop@euro-index.nl
www.euro-index.nl

