

EVCA210/EVCA210-UK

Electric vehicle charge-point adaptor



- Push Button for PE Pre-Test
- Push button for CP Error "E" simulation
- Push button for PE Error (Earth Fault) simulation
- Rotary switch providing PP State simulation
- Rotary Switch providing CP State simulation
- Type 1 to Type 2 Adaptor available (Supplied as standard with the EVCA210-UK)
- Carrying Case
- CAT II 300V Rating
- Comply with Low Voltage Directive LVD 2014/35

DESCRIPTION

The Megger EVCA210/EVCA210-UK are compact, simple to use adapters designed to perform all the functions required by the electrical contractor to fully test Mode 3 AC Electric Vehicle Charge-points. Specially designed to comply with UK, European and other International wiring regulations and standards, the EVCA210/EVCA210-UK may be used on all single and three phase electric vehicle charge points, with appropriate connectors. They are designed to test the function and safety of a charging point. The adapters allow you to conduct tests using appropriate single or multifunction test instruments on EV charging points in accordance with IEC/EN 61851-1 and IEC/HD 60364-7-722. Charging points should be tested as part of the initial installation and repeated periodically.

The Megger EVCA210/EVCA210-UK test adaptors are designed to simulate the connection of an electrical vehicle to the charging point under test. Connection of the adapter enables the operator to trigger the charging process by selecting the appropriate Proximity Pilot (PP) and Control Pilot (CP) states.

Test instruments can be connected using either the front mounted mains socket or the 4mm connection ports L1, L2, L3, N, PE. 2 additional CP signal terminals give the operator the ability to measure the CP signal using an oscilloscope.

In addition, the adapters have a manual PE Pre-Test feature that allows the user to test for dangerous voltages present on the PE, prior to any other testing of the charge point. If this test fails, testing must cease as potentially dangerous voltages may be present on the PE and faults must be investigated and rectified before continuing. The adapters also integrate two further manual tests: CP Error – simulates an error on the control pilot circuit; and a PE Error – where a disconnection of the PE circuit is produced. Both test ensure correct disconnection of the output of the charge point. Before carrying out testing with this adaptor, it is recommended that the user familiarize themselves with the relevant standards:

IEC 61851-1:2017, Electric vehicle conductive charging system - Part 1: General requirements

EVCA210/EVCA210-UK

Electric vehicle charge-point adaptor

IEC 60364-7-722:2018 Low-voltage electrical installations

- Part 7-722: Requirements for special installations

or locations - Supplies for electric vehicles and any

documentation relating to the charging station itself.

All Megger EVCA210 variants are supplied with a fixed Type

2 connector plug. The EVCA210-UK is also supplied with

a Type 2 to Type 1 adaptor plug as standard - optional for

other variants.

	EVCA210	EVCA210-UK
Input voltage 250 V (single phase system) / 430 V (three phase system),	■	■
Push Button for PE Pre-Test	■	■
Rotary switch providing PP State simulation	■	■
Rotary Switch providing CP State simulation	■	■
Push button for CP Error "E" simulation	■	■
Push button for PE Error (Earth Fault) simulation	■	■
Measurements on live conductors (L1, L2, L3 and N) and on PE conductor	■	■
Mains Socket: EVCA210-UK - 13A UK socket EVCA210 - Schuko socket (CEE 7/3)	■	■
CP signal test, two 4 mm ports for connection to an oscilloscope	■	■
CAT II 300 V	■	■
Fixed Type 2 Plug for charging points with panel mounted socket outlet or fixed cable with vehicle connector.	■	■
Type 1 to Type 2 adaptor for charging points with fixed cable and vehicle connector – Supplied with EVCA210-UK as standard		■
Soft Carry Case	■	■

EVCA210/EVCA210-UK

Electric vehicle charge-point adaptor

SPECIFICATIONS

Input voltage	Up to 250 V (single phase system) / up to 430 V (three phase system)	Altitude above sea level.	2000 m max.
Input Frequency	50/60 Hz,	Dimensions (W x H x L)	225 mm x 145 mm x 62 mm (without connection cable and test cable)
PE Pre-Test	Yes – Push button	Weight	Approx: 780g
CP Error “E” simulation	Push button	CE directive	Low Voltage Directive LVD 2014/35/EU
PP Simulation	NC, 13 A, 20 A, 32 A, 63 A	Safety	IEC/EN 61010-1:2010 IEC/EN 61010-2-030:2010
CP States	A, B, C, D	Working temperature range	0 to +40 °C
CP Error “E”	Yes	Storage temperature range	-10 to +50 °C
PE Error (Earth fault)	Yes	Reference humidity range	10 to 60% relative humidity w/o condensation
Measuring Ports L1, L2, L3, N and PE	Max. 250/430 V, CAT II 300 V, max. 10 A	Working humidity range	10 to 85% relative humidity w/o condensation
Mains socket	Max. 250 V, CAT II 300 V, max. 10 A, Note: Do not load mains socket simultaneously with measuring ports!		
CP Signal Test Ports	Approx. +/-12 V, CAT 0 (under normal condition)		
Measurement category	CAT II 300 V		

ORDERING INFORMATION

Description	Part number	Description	Part number
EVCA210-UK (UK Mains Socket)	1012-732	EVCA210 (Schuko Mains Socket)	1013-317
Included accessories		Included accessories	
Type 1 to Type 2 Charging Plug Adaptor		EVCA210	
Fixed Type 2 Charging Plug		Fixed Type 2 Charging Plug	
EVCA210 Carry Case		Soft Carry Case	
Instruction Manual		Instruction Manual	
		Optional accessories for EVCA210 & EVCA210-UK	
		EVCA210 Soft Carry Case	1013-318
		Type 1 to Type 2 Plug Adaptor	1014-901
		EVCA210-CH EV adaptor	1013-832
		EVCA210-AU EV adaptor	1015-141

SALES OFFICE

Megger Limited
Archcliffe Road Dover
CT17 9EN England
T +44 (0) 1304 502101
E UKsales@megger.com

EVCA_DS_en_V06

www.megger.com
ISO 9001
The word 'Megger' is a registered trademark

Megger[®]

Diensten van EURO-INDEX

EURO-INDEX is fabrikant, importeur en distributeur van diverse A-merken op het gebied van test- en meetinstrumenten. Daarnaast leveren wij een groot aantal diensten om het gebruik van deze instrumenten in uw bedrijfsvoering te optimaliseren. Dit omvat uiteraard onderhoud, reparatie en kalibratie van de instrumenten, maar ook kennisdeling via EURO-INDEX Academy en verhuur van instrumenten.

Geautoriseerd Service Centrum

EURO-INDEX b.v. is van alle vertegenwoordigde merken een Geautoriseerd Service Centrum. Dit betekent dat uw instrumenten worden behandeld door technici die zijn opgeleid door de fabrikant en beschikken over de juiste gereedschappen en software. Er worden uitsluitend originele onderdelen toegepast en de garantie van uw instrument, evenals de certificering (ATEX, EN50379, etc.) blijven intact.

Kalibratielaboratorium

Het laboratorium in Nederland beschikt over een RvA accreditatie naar EN-ISO/IEC 17025. Deze accreditatie geldt voor grootheden, zoals gespecificeerd in de scope bij accreditatienummer K105. RvA kalibratiecertificaten zijn internationaal geaccepteerd en is gelijkwaardig aan BELAC.



Mobiele Service

Naast de vaste kalibratielaboratoria in Zaventem en Capelle aan den IJssel beschikken wij ook over een laboratorium op wielen met de naam "Mobiele Service". Dit biedt vertrouwde service en kwaliteit, bij u voor de deur!

KWS®

KWS® is een uniek servicesysteem voor uw meetinstrumenten met periodiek onderhoud en kalibratie tegen vaste, lage kosten. Via een gratis webportal (mijnkws.be) heeft u altijd en overal beschikking over uw kalibratiecertificaten.

Verhuur van meetinstrumenten

- Uitgebreid assortiment
- Nauwkeurigheid aantoonbaar door actueel kalibratiecertificaat
- Deskundig advies
- Complete levering inclusief accessoires

EURO-INDEX Academy

- Trainingen, seminars en workshops
- Demonstratie- en instructievideo's
- Application notes



Servicebalie



Onderhoud, reparatie en kalibratie



Trainingen en seminars



Mobiele Service

Wijzigingen voorbehouden EURO-INDEX® VL 23001



BELGIË
Leuvensesteenweg 607
1930 Zaventem
T: 02 - 757 92 44
F: 02 - 757 92 64
sales@euro-index.be
www.euro-index.be

NEDERLAND
Rivium 2e straat 12
2909 LG Capelle a/d IJssel
T: +31 - (0)10 - 2 888 000
F: +31 - (0)10 - 2 888 010
verkoop@euro-index.nl
www.euro-index.nl

