



Multitec[®] 545/540

Series

Type/model	• Multitec 545: 066 13
	• Multitec 540: 066 12

Certificates

Certificate	TÜV 07 ATEX 553353 X
	• II 2G Ex db eb ib IIB T4 Gb basic device without leather bag for: ◦ CH ₄ , C ₃ H ₈ , C ₄ H ₁₀ , C ₉ H ₂₀ , H ₂ S, CO
	• II 2G Ex db eb ib IIC T4 Gb basic device with leather bag for: ◦ CH ₄ , C ₃ H ₈ , C ₄ H ₁₀ , C ₉ H ₂₀ , H ₂ S, CO, H ₂

Device data

Dimensions (W x D x H)	approx. 148 x 57 x 205 mm (5.83 x 2.24 x 8.07 in) approx. 148 x 57 x 253 mm (5.83 x 2.24 x 9.96 in) with supporting bracket
Weight	approx. 1000 g (35 oz), depending on equipment

Features

Display	monochrome, 320 x 240 pixel
Buzzer	• frequency: 2,4 kHz • volume: 80 dB (A) / 1m (3.28 ft)
Signal light	red
Pump	• vacuum: > 250 mbar • volume flow: typically 50 l/h ±20 l/h • pump error (F100) depending on volume flow: ◦ ≤ 20 l/h F100 certain ◦ > 20 l/h – ≤ 35 l/h F100 possible
Interface	USB 2.0
Memory	8 MB
Control	• ON/OFF key • 3 function keys • jog dial
Sensors	• IR for flammable gases (CH ₄) • IR for CO ₂ optional: • EC for CO, H ₂ , H ₂ S, O ₂

Operating conditions*

Operating temperature	-20 – 40°C (-4 °F – 104 °F)
Humidity	5 – 90% r.h., non-condensing
Atmospheric pressure	800 – 1100 hPa
Pressure at gas inlet	-175 – 65 hPa
Protection rating	IP54
Position of use	any

*Optional sensors can affect the operating conditions of the device.

Storage conditions

Storage temperature	-25 – 60°C (-13 °F – 140 °F) temperatures above 40°C (104 °F) reduce the service life of the sensors
---------------------	---

Power supply

Power supply	4 cells, type Mignon AA, optionally: • rechargeable batteries: NiMH • disposable batteries: Alkaline
Operating time, typical	at least 6 h
Battery capacity	2000 mAh
Charging time	approx. 3 h (complete charge), depending on capacity
Charging temperature	0 – 35°C (32 to 95 °F)
Charging voltage	12 V DC (max. 1 A)

Data transmission

Communication	USB
---------------	-----

Gas types

Default	CH ₄ , CO ₂
Optional	CO, H ₂ , H ₂ S, O ₂

Sensors

Note:

when using probes, the specified response times are longer.

Note for EC sensors:

at temperatures below 0 °C (32 °F) the specified response times and decay times may be longer.

Methane CH₄ (Interspace application)

Type	infrared sensor (IR)	
Use	Multitec 545/540	
Measuring range	0 ppm – 1.00% vol.	
Indication range	0 ppm – 2.5% vol.	
Resolution	• 50 ppm (0 – 950 ppm) • 0.01% vol (0.10 – 2.50% vol.)	
Response times	t ₅₀ < 9 s	t ₉₀ < 17 s
Warm-up time	< 60 s	
Temperature range	-20 – 40 °C (-4 °F – 104 °F)	
Measuring error	±15% of measured value (linearity), at least ±100 ppm	
Interference, known	all hydrocarbons C _x H _y	
Lifetime, expected	5 years	
Adjustment	test gas concentration: • zero point: hydrocarbon-free, clean air • CH₄: 1% vol., utilisable 0.50 – 1.00% vol.	

Methane CH₄ (Gas measuring application)

Type	infrared sensor (IR)	
Use	Multitec 545/540	
Measuring range	0 – 100% vol.	
Resolution	• 0.1% vol. (0 – 79.9% vol.) • 1% vol. (80 – 100% vol.)	
Response times	t ₅₀ < 9 s	t ₉₀ < 17 s
Warm-up time	< 30 s	
Temperature range	-20 – 40 °C (-4 °F – 104 °F)	
Measuring error	±1.5% of measured value, at least ±0.5% vol.	
Interference, known	all hydrocarbons C _x H _y	
Lifetime, expected	5 years	
Adjustment	test gas concentration: • zero point: hydrocarbon-free, clean air • CH₄: ◦ 100% vol. ◦ gas mixture 60% vol. CH₄ / 40% vol. CO₂	

Carbon dioxide CO₂

Type	infrared sensor (IR)
Use	Multitec 545/540
Measuring range	0 – 100% vol.
Resolution	• 0.1% vol. (0 – 9.9% vol.) • 1% vol. (10 – 100% vol.)
Response times	t ₉₀ < 20 s
Warm-up time	< 30 s
Temperature range	-20 – 40 °C (-4 °F – 104 °F)
Measuring error	±1.5% vol.
Zero point deviation	0.04% vol.
Interference, known	none
Humidity	5 – 90% r.h., non-condensing • short term: 0% r.h
Lifetime, expected	5 years
Adjustment	test gas concentration: • zero point: carbon dioxide-free, clean air • CO₂: ◦ 100% vol. ◦ gas mixture 60% vol. CH₄ / 40% vol. CO₂

Oxygen O₂

Type	electrochemical sensor (EC)
Use	Multitec 545/540
Measuring range	0 – 25% vol.
Indication range	0 – 30% vol.
Resolution	0.1% vol.
Response times	t ₉₀ < 30 s
Warm-up time	up to 90 s
Temperature range	-20 – 40 °C (-4 °F – 104 °F)
Measuring error	• linearity: ≤ 1.5% of measured value, minimum ≤ 0.3% vol. • long-term stability: ≤ 0.2% vol. (3 months)
Drift	< 2% within 3 months
Interference, known	none
Humidity	5 – 90% r.h., non-condensing • short term: 0% r.h
Lifetime, expected	2 years
Adjustment	test gas concentration: • zero point: ◦ oxygen-free air ◦ 100% vol. N₂ ◦ 100% vol. CH₄ • O₂: 20.9% vol., e.g. clean air

Carbon monoxide CO

Type	electrochemical sensor (EC)	
Use	Multitec 545/540	
Measuring range	0 – 500 ppm • lower limit: ◦ 0 – 100 ppm: 4 ppm ◦ > 100 ppm: 11 ppm	
Indication range	0 – 600 ppm	
Resolution	1 ppm	
Response times	t ₉₀ < 30 s	
Decay times	t ₁₀ ≤ 25 s	
Warm-up time	up to 90 s	
Temperature range	-20 – 40 °C (-4 °F – 104 °F)	
Measuring error	• ±3 %, minimum ±3 ppm (±3 digit) • long-term stability ◦ test gas: ≤ 4% of measured value ◦ zero point (fresh air): ≤ 1 ppm	
Drift	< 10% within 6 months	
Zero point deviation	• 0 – 100 ppm: 3 ppm • > 100 ppm: 13 ppm	
Interference, known	at 20°C (68 °F) • C ₂ H ₂ 100 ppm: approx. 90 ppm CO • C ₂ H ₄ 100 ppm: approx. 96 ppm CO • Cl ₂ 15 ppm: approx. 1 ppm CO • H ₂ 200 ppm: approx. 30 ppm CO • H ₂ S 50 ppm: approx. 1 ppm CO • NH ₃ 50 ppm: approx. 0 ppm CO • NO 50 ppm: approx. 15 ppm CO • SO ₂ 20 ppm: approx. 0 ppm CO	
Humidity	15 – 90% r.h., non-condensing • short term: 0% r.h	
Lifetime, expected	3 years	
Adjustment	test gas concentration: • zero point: clean air • CO: 40 ppm, utilisable 10 – 150 ppm	

Hydrogen H2

Type	electrochemical sensor (EC)	
Use	Multitec 545/540	
Measuring range	0 – 1000 ppm	
Indication range	0 – 1200 ppm	
Resolution	5 ppm	
Response times	t ₉₀ < 60 s	
Decay times	t ₁₀ ≤ 60 s	
Warm-up time	up to 90 s	
Temperature range	-20 – 40 °C (-4 °F – 104 °F)	
Measuring error	±10 %, minimum ±15 ppm (±3 digit)	
Drift	< 10% within 6 months	
Zero point deviation	±10 ppm (±2 digit)	
Interference, known	at 20°C (68 °F) • CO 200 ppm approx. 150 ppm H2 • H2S 25 ppm approx. -0.5 ppm H2 • N2O 20 ppm approx. -1 ppm H2 • NO 50 ppm approx. 20 ppm H2 • S2O 20 ppm approx. -1 ppm H2	
Humidity	15 – 90% r.h., non-condensing • short term: 0% r.h	
Lifetime, expected	3 years	
Adjustment	test gas concentration: • zero point: clean air • H2: 1000 ppm, utilisable 100 – 1000 ppm	

Hydrogen sulphide H₂S

Type	electrochemical sensor (EC)	
Use	Multitec 545	
Measuring range	0 – 5000 ppm	
	• lower limit:	1 ppm
Indication range	0 – 6000 ppm	
Resolution	• 1 ppm (0 – 100 ppm) • 2 ppm (100 - 998 ppm) • 0.02% vol. / 200 ppm (0.10 - 0.5% vol.)	
Response times	t ₉₀ < 60 s	
Decay times	t ₁₀ < 90 s	
Warm-up time	up to 120 s	
Temperature range	-20 – 40 °C (-4 °F – 104 °F)	
Measuring error	• ±3% or ±3 ppm (±3 digits) • ±3 ppm (long-term stability)	
Drift	< 10% within 6 months	
Zero point deviation	2 ppm	
Interference, known	at 20°C (68 °F) • CO 200 ppm: approx. 2 ppm • SO₂ 20 ppm: approx. 3 ppm • NO₂ 200 ppm: approx. -30 ppm • H₂ 100 ppm: approx. 2 ppm	
Humidity	15 – 90% r.h., non-condensing • short term: 0% r.h	
Lifetime, expected	2 years	
Adjustment	test gas concentration: • zero point: clean air • H₂S: 180 ppm, utilisable 10 – 1200 ppm	

Hydrogen sulphide H₂S

Type	electrochemical sensor (EC)	
Use	Multitec 540	
Measuring range	0 – 2000 ppm	
	• lower limit:	1 ppm
Indication range	0 – 2400 ppm	
Resolution	• 1 ppm (0 – 100 ppm) • 2 ppm (100 - 998 ppm) • 0.02% vol. / 200 ppm (0.10 - 0.2% vol.)	
Response times	t ₉₀ < 60 s	
Decay times	t ₁₀ < 90 s	
Warm-up time	up to 120 s	
Temperature range	-20 – 40 °C (-4 °F – 104 °F)	
Measuring error	• ±3% or ±3 ppm (±3 digits) • ±3 ppm (long-term stability)	
Drift	< 10% within 6 months	
Zero point deviation	2 ppm	
Interference, known	at 20°C (68 °F) • H₂ 2% vol.: approx. 150 ppm H₂S • Isopropanol 1% vol.: approx. 0 ppm H₂S • NH₃ 1000 ppm: approx. 0 ppm H₂S	
Humidity	15 – 90% r.h., non-condensing • short term: 0% r.h	
Lifetime, expected	2 years	
Adjustment	test gas concentration: • zero point: clean air • H₂S: 180 ppm, utilisable 10 – 1200 ppm	

Diensten van EURO-INDEX


EURO-INDEX is fabrikant van BLAUWE LIJN en importeur/distributeur van diverse A-merken test- en meet-instrumenten. Daarnaast leveren wij een groot aantal diensten om het gebruik van deze instrumenten in uw bedrijfsvoering te optimaliseren. Dit omvat onderhoud, reparatie en kalibratie van de instrumenten, maar ook kennisdeling via EURO-INDEX Academy en verhuur van instrumenten.

Geautoriseerd Service Centrum

EURO-INDEX is van de meeste merken in het assortiment een Geautoriseerd Service Centrum. Dit betekent dat uw instrumenten worden behandeld door technici die zijn opgeleid door de fabrikant en beschikken over de juiste gereedschappen en software. Er worden uitsluitend originele onderdelen toegepast en de garantie van uw instrument blijft intact, net als de certificering (ATEX, EN50379, etc.).

Kalibratielaboratorium

Ons service- en kalibratielaboratorium is BELAC-geaccrediteerd volgens de norm EN-ISO/IEC 17025. De accreditatie geldt voor grootheden en meetbereiken zoals vermeld in de gewaarmerkte bijlage met certificaatnummer [760-CAL](#).



KWS®

KWS® is een kalibratieformule voor uw test- en meetinstrumenten met periodiek onderhoud en kalibratie tegen vaste, lage kosten. Uw kalibratiecertificaten zijn digitaal beschikbaar via Mijn KWS (gratis webportaal en app) en door de QR-code te scannen op de kalibratiesticker van het instrument.

Verhuur van meetinstrumenten

Er zijn diverse situaties waarbij huren handig is:

- U heeft tijdelijk extra toestellen nodig.
- Uw eigen meetinstrument wordt onderhouden en/of gekalibreerd.
- U moet een eenmalige meting verrichten.

EURO-INDEX Academy

- Trainingen, infosessies en workshops
- Demonstratie- en instructievideo's
- Application notes



Servicebalie



Onderhoud, reparatie en kalibratie



Trainingen en seminars



Kalibratielaboratorium

Wijzigingen voorbehouden EURO-INDEX® VL 26001



BELGIË
Leuvensesteenweg 607
1930 Zaventem
T: 02 - 757 92 44
sales@euro-index.be
www.euro-index.be

NEDERLAND
Rivium 2e straat 12
2909 LG Capelle a/d IJssel
T: +31 - (0)10 - 2 888 000
verkoop@euro-index.nl
www.euro-index.nl

